TSK-GEL® STAT Series Anion Exchange Products

Part Numbers: 21960, TSKgel Q-STAT, 3mm ID X 3.5cm, $10\mu m$

21961, TSKgel Q-STAT, 4.6mm ID X 10cm, 7µm

21962, TSKgel DNA-STAT, 4.6mm ID X 10cm, 5µm

This sheet contains the recommended operating conditions and the specifications for TSK-GEL STAT Series Anion Exchange columns. Installation instructions and column care information are described in a separate Instruction Manual.

OPERATING CONDITIONS

1. Shipping Solvent: Ion-Exchanged Water

1.0 - 2.0mL/min (P/N 21960) 0.5 - 1.4mL/min (P/N 21961) Standard Flow Rate: 0.3 - 0.6mL/min (P/N 21962)

When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the

maximum pressure drop.

10Mpa (P/N 21960)

10Mpa (P/N 21961) 15Mpa (P/N 21962) Max. Pressure:

pH Range: 3.0 - 10.0 (pH above 10 can only be used for a short time)

<50%. When solvent in column is replaced by distilled or ion-exchanged water, feed the solvent slowly, at flow Organic Conc.:

rates <0.5mL/min (Q-STAT) and <0.25mL/min (DNA-STAT).

Temperature:

Adsorbed materials can be stripped from the column by repeated injection with one of the following cleaning Cleaning Solvents:

(1) 0.1mol/L NaOH, or

(2) 20 ~ 40% Acetic acid, or

(3) Solution containing aqueous organic solvent such as methanol or acetonitrile, or

(4) Solution containing a solubilizer such as urea and non-ionic surfactants

Short term storage: keep the column filled with low ionic strength eluent.

Storage: For long term storage, replace the solvent in the column with distilled or ion-exchanged water, at flow rates

<0.5mL/min (Q-STAT) and <0.25mL/min (DNA-STAT).

SPECIFICATIONS

The performance of TSK-GEL Q-STAT and TSK-GEL DNA-STAT columns are tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:

≥ 200 (PN 21960) **Number of Theoretical Plates** \geq 4,000 (PN 21961)

(N):

≥ 4,000 (PN 21962)

0.8 - 1.8 (PN 21960) 1.0 - 2.0 (PN 21961) 2. Asymmetry Factor (AF):

1.0 - 2.0 (PN 21962)